



CERTIFICATE OF ANALYSIS

Product Name	Mueller Hinton Broth (5 ml)
Catalogue No	01018
Batch No	2506065
Dehydrated Brand / Batch No / Expiry Date	Oxoid / 3517048 / 06-2027
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	27.06.2025
Expiry Date	27.12.2025
Storage Conditions	2 – 12 °C

Physical & Chemical Specifications

Appearance	Clear	Appropriate
Color	Pale yellow	Appropriate
pH	7,4 ± 0,2	7,42

Microbiological Specifications: Productivity and acc. To ISO/TS 11133

Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)		Incubation: 35±2 °C:24-72 h		
Sterility	Incubation: (20-25°C :3 d, 30-35°C :2 d)		Sterile		
QC Strains	ATCC	Results			
		Control	Product	Growth	
Staphylococcus aureus	25923	Good	Good	Good	Appropriate
Escherichia coli	25922	Good	Good	Good	Appropriate
Pseudomonas aeruginosa	27853	Good	Good	Good	Appropriate
Enterococcus faecalis	29212	Good	Good	Good	Appropriate

QC Date	04.07.2025
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net