



CERTIFICATE OF ANALYSIS

Product Name	Urea Agar (5 ml)					
Catalogue No	01041					
Batch No	2508089					
Dehydrated Brand / Batch No / Expiry Date	Urea Agar Base: Condalab/405043/05-2027					
Neutralizer Brand / Batch No / Expiry Date	-					
Production Date	21.08.2025					
Expiry Date	21.02.2026					
Storage Conditions	2 - 12° C					
Physical & Chemical Specifications						
Appearance	Must have a smooth surface			Appropriate		
Color	Orange-yellow			Appropriate		
pH	7,0 ± 0.2			7,06		
Microbiological Specifications: Productivity and acc. To ISO/TS 11133						
Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100-1000 cfu)			Incubation: 35±2 °C:18-24 h		
Sterility	Incubation: (20-25°C : 3 d, 30-35°C : 2 d)			Sterile		
QC Strains	ATCC	Results				
		Control	Product	Ureaz Reaction	Colour	Results
Escherichia coli	25922	Good	Good	-	Yellow	Appropriate
Proteus mirabilis	43071	Good	Good	+	Pink	Appropriate

QC Date	28.08.2025
QC Manager	Ali Aras

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net