



CERTIFICATE OF ANALYSIS

Product Name	Cetrimide Agar
Catalogue No	02008
Batch No	2303007
Dehydrated Brand / Batch No / Expiry Date	Cetrimide Agar: Oxoid/ 3450916/ 03-2025
Neutralizer Brand / Batch No / Expiry Date	Glycerol: Merck /W252506 / 06.2024
Production Date	02.03.2023
Expiry Date	02.07.2023
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Light amber, opalescent, with precipitate	Appropriate
pH	7,2 ± 0,2	7,24

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 30-35 °C: 18-24h	
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Pseudomonas aeruginosa	9027	73	72	Green - Yellow	Appropriate
Pseudomonas aeruginosa	27853	74	73	Green - Brown	Appropriate
Escherichia coli	25922	Partial inhibition	Partial inhibition	Partial inhibition	Appropriate
Straptococcus aureus	25923	inhibition	inhibition	inhibition	Appropriate

Acceptance Criteria: %50-200

QC Date	09.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net