



CERTIFICATE OF ANALYSIS

Product Name	Chocolate Agar				
Catalogue No	02009				
Batch No	2303045				
Dehydrated Brand / Batch No / Expiry Date	CBA: Oxoid/ 3554393/ 10-2027				
Neutralizer Brand / Batch No / Expiry Date	-				
Production Date	14.03.2023				
Expiry Date	14.08.2023				
Storage Conditions	2-12°C				
Physical & Chemical Specifications					
Appearance	Must have a smooth surface			Appropriate	
Color	Dark brown			Appropriate	
pH	7,2 ± 0,2			7,26	
Microbiological Specifications					
Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 30-35 °C: 24-48h	
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Haemophilus influenzae	10211	72	73	White small colonies	Appropriate
Neisseria gonorrhoeae	49226	73	75	White small colonies	Appropriate
Staphylococcus pyogenes	19615	74	76	Cream colonies	Appropriate
Staphylococcus pneumoniae	6305	74	75	Cream colonies	Appropriate
Acceptance Criteria: %50-200					
QC Date	21.03.2023				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net