



CERTIFICATE OF ANALYSIS

Product Name	Fluid Thioglycollate Medium (S.T.B) (100mL)
Catalogue No	01204
Batch No	2303049
Dehydrated Brand / Batch No / Expiry Date	Thioglycollate Medium: Himedia/ 0000518698 / 12-2026
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	14.03.2023
Expiry Date	14.09.2023
Storage Conditions	2 - 25°C

Physical & Chemical Specifications

Appearance	Clear	Appropriate
Color	Yellow, beige	Appropriate
pH	7,1 ± 0,2	7,12

Microbiological Specifications

Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 30-35 °C: 24h-14d	
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Bacillus subtilis	6633	Good	Good	Good	Appropriate
Staphylococcus aureus	6538	Good	Good	Good	Appropriate
Candida albicans	10231	Good	Good	Good	Appropriate
Pseudomonas aeruginosa	9027	Good	Good	Good	Appropriate
Aspergillus brasiliensis	16404	Good	Good	Good	Appropriate
Clostridium sporogenes	19404	Good	Good	Good	Appropriate

QC Date	21.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net