



CERTIFICATE OF ANALYSIS

Product Name	Rodak Sabouraud Dextrose Agar
Catalogue No	05006
Batch No	2303068
Dehydrated Brand / Batch No / Expiry Date	SDA: Condalab/312191/ 12-2026
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	22.03.2023
Expiry Date	22.08.2023
Storage Conditions	2-25°C

Physical & Chemical Specifications

Appearance	Convex, smooth	Appropriate
Color	Amber	Appropriate
pH	5,6 ± 0,2	5,62

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)	Incubation: 25±2 °C:48h-5d
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)	Sterile
QC Strains	ATCC	Results
		ControlProductGrowth
Candida albicans	10231	7375GoodAppropriate
Aspergillus brasiliensis	16404	7274GoodAppropriate

Acceptance Criteria: %50-200

QC Date	29.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net