



CERTIFICATE OF ANALYSIS

Product Name	HEKTOEN ENTERIC AGAR
Catalogue No	02028
Batch No	2303070
Dehydrated Brand / Batch No / Expiry Date	HEKOEN ENTERIC AGAR: Condalab/ 307052 / 07-2026
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	22.03.2023
Expiry Date	22.06.2023
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Dark green	Appropriate
pH	7,5 ± 0,2	7,52

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 30-35 °C: 24-48h	
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Proteus mirabilis	43071	74	75	Good	Appropriate
Salmonella typhimurium	14028	72	74	Good	Appropriate
Shigella flexneri	12022	73	75	Good	Appropriate
Enterococcus faecalis	29212	Partial Inhibition	Partial Inhibition	Partial Inhibition	Appropriate
Escherichia coli	25922	Partial Inhibition	Partial Inhibition	Partial Inhibition	Appropriate

Acceptance Criteria: %50-200

QC Date	29.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net