



CERTIFICATE OF ANALYSIS

Product Name	Mueller Hinton Agar				
Catalogue No	02033				
Batch No	2303074				
Dehydrated Brand / Batch No / Expiry Date	MHA: Condalab/306031/ 06-2026				
Neutralizer Brand / Batch No / Expiry Date	-				
Production Date	23.03.2023				
Expiry Date	23.09.2023				
Storage Conditions	2-12°C				
Physical & Chemical Specifications					
Appearance	Must have a smooth surface			Appropriate	
Color	Light amber			Appropriate	
pH	7,3 ± 0,2			7,35	
Microbiological Specifications					
Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 35±2 °C:24-48h	
Sterility	Incubation: (20-25°C : 3 d, 30-35°C : 2 d)			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Escherichia coli	25922	74	75	Good	Appropriate
Staphylococcus aureus	25923	74	75	Good	Appropriate
Pseudomanas aeruginosa	27853	73	74	Good	Appropriate
Enterococcus faecalis	29212	73	74	Good	Appropriate
QC Date	30.03.2023				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net