



CERTIFICATE OF ANALYSIS

Product Name	HEKTOEN ENTERIC AGAR
Catalogue No	02028
Batch No	2311024
Dehydrated Brand / Batch No / Expiry Date	Hektoen Enteric Agar: Condalab/ 402102 / 02-2027
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	04.11.2023
Expiry Date	04.02.2024
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Yellowish green / Reddish purple	Appropriate
pH	7,5 ± 0,2	7,51

Microbiological Specifications: Productivity and acc. To ISO/TS 11133

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)	Incubation: 35±2 °C: 24-48h
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)	Sterile
QC Strains	ATCC	Results
		ControlProductGrowth
Proteus mirabilis	43071	7373Black coloniesAppropriate
Salmonella typhimurium	14028	7275Black coloniesAppropriate
Shigella flexneri	12022	7173Green-Blue ColoniesAppropriate
Enterococcus faecalis	29212	InhibitionInhibitionInhibitionAppropriate
Escherichia coli	25922	Partial InhibitionPartial InhibitionOrange coloniesAppropriate

Acceptance Criteria: %50-200

QC Date	11.11.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net