



CERTIFICATE OF ANALYSIS

Product Name	Chromogenic VRE Agar				
Catalogue No	02112				
Batch No	2312008				
Dehydrated Brand / Batch No / Expiry Date	Chromogenic VRE Agar: Chromagar/ P003242/03-2025				
Neutralizer Brand / Batch No / Expiry Date	Chromogenic VRE Supplement: Chromagar/ P003194/ 05-2026				
Production Date	04.12.2023				
Expiry Date	19.02.2024				
Storage Conditions	2-12°C				
Physical & Chemical Specifications					
Appearance	Must have a smooth surface			Appropriate	
Color	White			Appropriate	
pH	6,8±0,2			6,82	
Microbiological Specifications: Productivity and acc. To ISO/TS 11133					
Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 35±2 °C:24 h	
Sterility	Incubation: (30-35°C : 2 d)			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Enterococcus faecalis	51299	73	74	Pink to mauve	Appropriate
Enterococcus faecalis	29212	Inhibition	Inhibition	Inhibition	Appropriate
Escherichia coli	25922	Inhibition	Inhibition	Inhibition	Appropriate
Acceptance Criteria: %50-200					
QC Date	08.12.2023				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net