



CERTIFICATE OF ANALYSIS

Product Name	DNase Agar
Catalogue No	02021
Batch No	2312014
Dehydrated Brand / Batch No / Expiry Date	DNase Test Agar: Condalab/207022/07-2025
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	05.12.2023
Expiry Date	05.03.2024
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Trace to slightly hazy	Appropriate
Color	Light amber	Appropriate
pH	7,3±0,2	7,35

Microbiological Specifications: Productivity and acc. To ISO/TS 11133

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)	Incubation: 35± 2°C:24 h
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)	Sterile
QC Strains	ATCC	Results
		ControlProductGrowth
Staphylococcus aureus	25923	GoodGoodDNase PozitiveAppropriate
Escherichia coli	25922	GoodGoodDNase NegativeAppropriate

Acceptance Crtieria: %50-200

QC Date	08.12.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net