



CERTIFICATE OF ANALYSIS

Product Name	Mueller Hinton Agar w / % 5 Horse Blood + Beta Nad				
Catalogue No	02035				
Batch No	2312021				
Dehydrated Brand / Batch No / Expiry Date	MHA: Condalab/401191/ 01-2027				
Neutralizer Brand / Batch No / Expiry Date	BETA-NAD / SLCL5420 / 01-2025				
Production Date	06.12.2023				
Expiry Date	06.03.2024				
Storage Conditions	2-12°C				
Physical & Chemical Specifications					
Appearance	Must have a smooth surface			Appropriate	
Color	Light amber			Appropriate	
pH	7,3 ± 0,2			7,36	
Microbiological Specifications: Productivity and acc. To ISO/TS 11133					
Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 35±2 °C:24-48h	
Sterility	Incubation: (20-25°C : 3 d, 30-35°C : 2 d)			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Haemophilus influenzae	49766	73	73	Good	Appropriate
Streptococcus pneumoniae	6305	72	74	Good	Appropriate
Campylobacter jejuni	33291	71	74	Good	Appropriate
Acceptance Criteria:%50-200					
QC Date	11.12.2023				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net