



CERTIFICATE OF ANALYSIS

Product Name	%5 Sheep Blood Agar
Catalogue No	02001
Batch No	2407084
Dehydrated Brand / Batch No / Expiry Date	CBA: Oxoid/3749512/11-2028
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	24.07.2024
Expiry Date	24.10.2024
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Red	Appropriate
pH	7,3 ± 0,2	7,32

Microbiological Specifications: Productivity and acc. To ISO/TS 11133

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)	Incubation: 30-35 °C: 18-24h
Sterility	Incubation: (30-35°C : 4d)	Sterile
QC Strains	ATCC	Results
		ControlProductGrowth
Escherichia coli	25922	7273Good (γ hemolysis)Appropriate
Staphylococcus aureus	25923	7375Good (β hemolysis)Appropriate
Streptococcus pyogenes	19615	7373Good (β hemolysis)Appropriate
Streptococcus pneumoniae	6305	7474Good (α hemolysis)Appropriate

Acceptance Criteria: %50-200

QC Date	31.07.2024
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net