



## CERTIFICATE OF ANALYSIS

|                                            |                            |
|--------------------------------------------|----------------------------|
| Product Name                               | %5 Sheep Blood Agar        |
| Catalogue No                               | 02001                      |
| Batch No                                   | 2410043                    |
| Dehydrated Brand / Batch No / Expiry Date  | CBA: Oxoid/3749512/11-2028 |
| Neutralizer Brand / Batch No / Expiry Date | -                          |
| Production Date                            | 14.10.2024                 |
| Expiry Date                                | 14.01.2025                 |
| Storage Conditions                         | 2-12°C                     |

### Physical & Chemical Specifications

|            |                            |             |
|------------|----------------------------|-------------|
| Appearance | Must have a smooth surface | Appropriate |
| Color      | Red                        | Appropriate |
| pH         | 7,3 ± 0,2                  | 7,33        |

### Microbiological Specifications: Productivity and acc. To ISO/TS 11133

|                                    |                                                            |                                   |
|------------------------------------|------------------------------------------------------------|-----------------------------------|
| Method of control: Counting Method | Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu) | Incubation: 30-35 °C: 18-24h      |
| Sterility                          | Incubation: (30-35°C : 4d)                                 | Sterile                           |
| QC Strains                         | ATCC                                                       | Results                           |
|                                    |                                                            | ControlProductGrowth              |
| Escherichia coli                   | 25922                                                      | 7274Good (γ hemolysis)Appropriate |
| Staphylococcus aureus              | 25923                                                      | 7374Good (β hemolysis)Appropriate |
| Streptococcus pyogenes             | 19615                                                      | 7474Good (β hemolysis)Appropriate |
| Streptococcus pneumoniae           | 6305                                                       | 7375Good (α hemolysis)Appropriate |

Acceptance Criteria: %50-200

|            |            |
|------------|------------|
| QC Date    | 21.10.2024 |
| QC Manager | Ali ARAS   |

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.  
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY  
Telefon:0216 354 90 54 E-mail: [info@bio-cell.net](mailto:info@bio-cell.net)