



CERTIFICATE OF ANALYSIS

Product Name	TSB STERILITY TEST BROTH (100ml)				
Catalogue No	01207				
Batch No	2411067				
Dehydrated Brand / Batch No / Expiry Date	Tryptic Soy Broth : HIMEDIA/E0006522095/05-2029				
Neutralizer Brand / Batch No / Expiry Date	-				
Production Date	19.11.2024				
Expiry Date	19.05.2025				
Storage Conditions	2-25°C				
Physical & Chemical Specifications					
Appearance	Clear		Appropriate		
Color	Yellow		Appropriate		
pH	7,3 ± 0,2		7,36		
Microbiological Specifications: Productivity and acc. To ISO/TS 11133					
Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)		Incubation: 35±2 °C:24-48 h 25±2 °C:6 d		
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Bacillus subtilis	6633	Good	Good	Good	Appropriate
Staphylococcus aureus	6538	Good	Good	Good	Appropriate
Candida albicans	10231	Good	Good	Good	Appropriate
Pseudomonas aeruginosa	9027	Good	Good	Good	Appropriate
Aspergillus brasiliensis	16404	Good	Good	Good	Appropriate
QC Date	26.11.2024				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net