



CERTIFICATE OF ANALYSIS

Product Name	Triple Sugar Iron Agar					
Catalogue No	01037					
Batch No	2501015					
Dehydrated Brand / Batch No / Expiry Date	TSI/207013 / 07-2025					
Neutralizer Brand / Batch No / Expiry Date	-					
Production Date	03.01.2025					
Expiry Date	03.07.2025					
Storage Conditions	2 - 12 ° C					
Physical & Chemical Specifications						
Appearance	Must have a smooth surface			Appropriate		
Color	Red			Appropriate		
pH	7.4 ± 0.2			7.42		
Microbiological Specifications: Productivity and acc. To ISO/TS 11133						
Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 35±2 °C:18-24 h		
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile		
QC Strains	ATCC	Results				
		Control	Product	Slant	Butt	H2S
Escherichia coli	25922	Good	Good	Acid	Acid	-
Salmonella typhimurium	14028	Good	Good	Alkaline	Acid	+
QC Date	10.01.2025					
QC Manager	Ali ARAS					

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net