



CERTIFICATE OF ANALYSIS

Product Name	Salmonella Shigella Agar
Catalogue No	02055
Batch No	2501018
Dehydrated Brand / Batch No / Expiry Date	SS AGAR: Oxoid / 3634887/10-2027
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	03.01.2025
Expiry Date	18.03.2025
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Red - orange	Appropriate
pH	7,0 ± 0,2	7,05

Microbiological Specifications: Productivity and acc. To ISO/TS 11133

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)	Incubation: 35±2 °C: 18-24h
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)	Sterile
QC Strains	ATCC	Results
		ControlProductGrowth
Salmonella typhimurium	14028	7377GoodAppropriate
Enterococcus faecalis	29212	Partial InhibitionPartial InhibitionPartial InhibitionAppropriate
Escherichia coli	25922	Partial InhibitionPartial InhibitionPartial InhibitionAppropriate

Acceptance Criteria: %50-200

QC Date	10.01.2025
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net