



CERTIFICATE OF ANALYSIS

Product Name	UTI Agar
Catalogue No	02083
Batch No	2507092
Dehydrated Brand / Batch No / Expiry Date	UTI Agar : Himedia/0000641830/03-2027
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	31.07.2025
Expiry Date	31.10.2025
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Light amber	Appropriate
pH	6,8±0,2	6,86

Microbiological Specifications: Productivity and acc. To ISO/TS 11133

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)		Incubation: 35±2 °C:24 h		
Sterility	Incubation: 30-35°C : 2d			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Enterococcus faecalis	29212	74	73	Blue, small	Appropriate
Escherichia coli	25922	72	72	Pink-purple	Appropriate
Staphylococcus aureus	25923	72	73	Blue to purple, mucoid	Appropriate
Proteus mirabilis	12453	73	74	Colorless	Appropriate
Klebsiella pneumoniae	4352	74	72	Light brown	Appropriate
Pseudomonas aeruginosa	9027	72	71	Golden yellow	Appropriate

Acceptance Criteria: %50-200

QC Date	07.08.2025
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net